

Finished digital thermometer factory inspection record

temperature: 24°C humidity: 32%

Lot No.	26YF1B0781	Model	MC-201	Order No.	
Inspection standard	GB/T 21416, ASTM E 1112-00, EN 80601-2-56, 浙械注准 20192070364	Batch	1500	Remark	The number of samples is greater than thirteen

Inspection method: sampling level: S-3 AQL:0.65 AC:RE % sampling quantity: 13

Inspection Item	Inspection requirement	Actual measurement/defect	Judgment OK/NG
accuracy	0.1°C/0.1°F	Pass	ok
repeatability	S ≤ 0.2°C	Pass	ok

Inspection equipment: Ultra high precision constant temperature bath: CGJDHWC-14 (32.5); CGJDHWC-02, (35), CGJDHWC-02 (37), CGJDHWC-15 (41.5), CGJDHW-10 (42.5)

Indicating value max error	Precision error requirement: ±0.1°C (35.0°C ~42.0°C) ±0.2°C (<35.0°C / >42.0°C)	Sample quantity													
		thermostatic water bath temperature	1	2	3	4	5	6	7	8	9	10	11	12	13
	32.5°C/90.5°F	32.5	32.5	32.5	32.5	32.5	32.5	32.5	32.5	32.6	32.5	32.5	32.5	32.5	32.5
	35°C/95°F	35.0	35.0	35.0	35.0	35.0	35.0	35.0	35.0	35.1	35.0	35.0	35.0	35.0	35.0
	37°C/98.6°F	37.0	37.0	37.0	37.0	37.0	37.0	37.0	37.0	37.1	37.0	37.0	37.0	37.0	37.0
	41.5°C/106.7°F	41.5	41.5	41.5	41.5	41.5	41.5	41.5	41.5	41.6	41.5	41.5	41.5	41.5	41.5
	42.5°C/108.5°F	42.5	42.5	42.5	42.5	42.5	42.5	42.5	42.5	42.6	42.5	42.5	42.5	42.5	42.5

Inspection method: sampling level: I AQL:1.0 AC:RE 1/2 package sampling quantity: 50

Inspection Item	Inspection requirement	Actual measurement/defect	Judgment OK/NG
exterior	Outer looking should be clean, there can't be scratch, glue and destroy on the body, there won't be plastic overlap and salient point on the LCD display, there won't be obvious wear trace on the shell.	Pass	ok
	Color of the body and cap should be same in one batch, there shouldn't be obvious color deviation.	Pass	
	Sensor top should be smooth, no burr on the edge.	Pass	
	Button should be flexible and firm.	Pass	
Logo, package, label	Files of the package, label, instrument, logo printing should meet customer's requirement.	Pass	ok
	Printing of the box, instrument and label should be clear, correct, papers should be clean.	Pass	ok
	Box bond place should be smooth and firm.	Pass	ok

Inspection method: sampling level: S-3 AQL:0.65 AC:RE % function indicate sampling quantity required: 13


Inspection Item	Inspection requirement	Actual measurement/defect	Judgment OK/NG
Function indicating requirem	Self inspection and memory function: when the thermometer is turned on, the LCD display of the thermometer should be clear, without any defects such as blurring, missing stroke, stroke deformation, etc., and then the last measured value should be displayed.	Pass	ok
	Buzzing prompt function: When the thermometer reaches a stable measurement value, it should give a prompt and a beeping prompt voice.	Pass	ok
	Automatic shutdown function: thermometer shall be automatically shut down within 8~10 minutes after measurement.	Pass	ok



**Finished digital thermometer factory inspection record**

Function indicating requirement	High and low temperature display function: when the measured temperature of the thermometer is lower than 32 °C, the LCD will display "LO °C". When the measured temperature is higher than 42.9 °C, the LCD will display "Hi °C".	Pass	ok
	Low voltage prompt function: When the voltage of the thermometer is lower than 90% (or meets the low voltage value of 1.25 V proposed by the manufacturer, but should not be higher than 90% of the rated value), the low voltage prompt should appear	Pass	ok
	Indication range: 32°C ~ 42.9°C, 90°F ~ 109.9°F	Pass	ok
	The shell of the thermometer shall be able to withstand biological and physical cleaning without functional degradation	Pass	ok
	High temperature alarm: when the temperature measured by the thermometer is $\geq 37.8$ °C, the thermometer will emit a warning voice with high frequency.	Pass	ok
	Measurement time: Normal type $\leq 50S$ Fast type $\leq 30S$	Pass	ok

Inspection method: sampling level: S-2 AQL:0.65 AC:RE% function indicate sampling quantity required: 8

Inspection Item	Inspection requirement	Actual measurement/defect	judgment OK/NG
Inspection equipment	Medical leakage current tester: YYLDLCSY-02; Programmable voltage withstand insulation tester: CKNYJCSY-01		
Withstand voltage test	Basic voltage DC1.5V Test voltage AC1500V No breakdown and flashover after 1min	Pass	ok
Leakage current test	Continuous leakage current at normal operating temperature under AC/DC:	1 2 3 4 5 6 7	
		0.00mA 0.00mA 0.00mA 0.00mA 0.00mA 0.00mA 0.00mA	
	Shell leakage current under normal condition $\leq 0.1mA$	8 9 10 11 12 13 14	
		0.00mA	
	Continuous leakage current at normal operating temperature under AC/DC:	1 2 3 4 5 6 7	
		0.00mA 0.00mA 0.00mA 0.00mA 0.00mA 0.00mA 0.00mA	
	Auxiliary leakage current of patients in normal state $\leq 0.01mA$	8 9 10 11 12 13 14	
		0.00mA	
	Continuous leakage current at normal operating temperature under AC/DC:	1 2 3 4 5 6 7	
		0.00mA 0.00mA 0.00mA 0.00mA 0.00mA 0.00mA 0.00mA	
	Shell leakage current under single fault condition $\leq 0.5mA$	8 9 10 11 12 13 14	
		0.00mA	
Continuous leakage current at normal operating temperature under AC/DC:	1 2 3 4 5 6 7		
	0.00mA 0.00mA 0.00mA 0.00mA 0.00mA 0.00mA 0.00mA		
Auxiliary leakage current under single fault state $\leq 0.05mA$	8 9 10 11 12 13 14		
	0.00mA		
Judgment:	Pass <input checked="" type="checkbox"/>	Default <input type="checkbox"/>	ok

Inspector/Date: Bai-Haixia  
Feb. 3. 2026

Auditor/Date: J. Wang  
Feb. 3. 2026.