


# CERTIFICATE OF ANALYSIS

PRODUCT NAME:	Three Function Hospital Bed	Lot. NO.	20260130	QTY	50	Test QTY	50	Standard	GB/T14710-2009
Model NO.	SKD-C	Inspector	Jian Tang	MFG Date	2026/1/30	Checker	Wu Weibao	DATE	2026/2/4
NO	Name of item	Technology specification						Test Result	Conclusion
1	Appearance	Appearance should be smooth, without irregularity , scars, distortion or deflection .						Compliant	Pass
2	Dimension	Dimension: 2200mm×900mm×500mm, Tolerance: ±10%. Height adjustment: 500-750mm, Tolerance: ±10%. Lifting speed: >=300mm/min.						Compliant	Pass
3	Movement range	Back-rest lifting: 80°, Tolerance: ±10%. Leg-rest lifting: 45°, Tolerance: ±10%. Height adjust from 500-750mm, Tolerance: ±10%.						Compliant	Pass
4	Capacity	Electric bed should bear 200kg average loading, Hi-Low of bed surface, backrest and leg rest lifting and height adjustable should be stable without obvious shaking, noise and other unusual performance						Compliant	Pass
5	Motor	all functions are adjustable by motor are in good condition						Compliant	Pass
Test result	<p>According to the test, all qualification reach the manufacture's standard request GB/T14710-2009</p> <p>Test result</p> <p>Quality managing department: Jian Tang Date:Feb.6.2026</p> <div style="text-align: right; margin-top: 20px;">  </div>								