

FOSHAN FEIYANG MEDICAL EQUIPMENT CO., LTD.
CERTIFICATE OF ANALYSIS

NO: FSYF-COA20250823002D

Product Name	COMMODE CHAIR / SILLA CON INODORO	Lot.No.	FY20250823	QTY	700PCS	Test QTY	700PCS	Standard	GB/T 24434 - 2009 (ISO13485: 2016)
Model No.	FY894	Inspector	Liu zhenyan	MFG Date	Aug.23,2025	checker	PENG LI	DATE	Aug.23,2025
NO.	Name of item	Technology specification						Result	Single conclusion
1	Appearance structure	Appearance should be smooth, without irregularity , scars, distortion or deflection .						Compliant	Pass
2	Dimension	Base size of commode should match the requirement of Form 1, limit deflection is no more than ± 5 cm.						Compliant	Pass
3	Stability	Four legs should be stable on the floor without swing.						Compliant	Pass
4	Leak resistant	Commode bucket and urinal should have well leak resistant, without water leakage.						Compliant	Pass
5	Colour transfer	Plastic parts should have no colour transfer.						Compliant	Pass
Test result	According to the test, all qualification reach the manufacture's standard request GB/T 24434 - 2009 (ISO13485: 2016) . Quality managing department: PAN QIU WEI Date:Aug.23,2025								

