FOSHAN FEIYANG MEDICAL EQUIPMENT CO., LTD. <u>CERTIFICATE OF ANALYSIS</u>

NO: FSFY-COA20230526001A

Product Name		STICK / BASTON		Lot.No.	FY20230525	QTY	3000PCS	Test QTY	3000PCS	Standard	GB/T 19545.1-2012 (ISO13485: 2016)
Model No.		FY920L		Inspector	ZOU MENG XIA	MFG Date	MAY 26,2023	checker	PENG LI	DATE	MAY 26,2023
NO.	Name of item		Technology specification							Result	Single conclusion
1	Appearance structure requirement		Appearance should be smooth, without irregularity, scars, distortion or deflection.							Compliant	Pass
2	Dimension		Base size of commode should match the requirement of Form 1, limit deflection is no more than ± 5 cm.							Compliant	Pass
3	Stability		The leg should be stable on the floor without swing.							Compliant	Pass
4	Colour transfer		Plastic parts should have no colour transfer.							Compliant	Pass
Test result According to the test, all qualification reach the manufacture's standard request GB/T 19545.1-2012 (ISO13485: 2016)										/4 //	
168	uit							Quality 1	nanaging departmen	rt: PAN QIU V	VEI Date:MAY 26,2023