

JIANGSU ZHENGKANG MEDICAL APPARATUS CO., LTD.
CERTIFICATE OF ANALYSIS

Product: Syringe : 1ml, 5ml, 10ml, 50ml
MFG. Date:11.2024
Inspection Basis: ISO7886-1:2017

Lot No.: 241120
EXP. Date:11.2029
Quantity:1002800pcs

Inspection item	Technical demands	Testing result
Sterility	Sterility	Sterility
Pyrogenicity	Non-Pyrogenicity	Non-Pyrogenicity
Hemolysis	Hemolysis rate<5%	1.2%
Toxicity	Non-toxicity	Non-toxicity
Reducing matter	The total amount of potassium permanganate solution used [c(KmnO4)=0.002mol/l] shall not exceed 0.5ml	0.1ml
Metal ions	The intensity of the color produced in the test solution shall not be deeper than that of the standard matching solution.	Qualified
Titration acidity	No more than 1ml of either standard volumetric solution shall be required for the indicator to change to the color grey.	0.2ml
Subcutaneous stimulate	No partial subcutaneous stimulate reaction	Qualified
Color lustre	Clear and colorless	Qualified
Capacity test	1ml:±4% of expelled volume, 10ml:±4% of expelled volume, 50ml:±4% of expelled volume	Qualified
Conclusion: Comply with the standard of ISO8536-1:2004		



Issuing Date: DEC.25.2024 Inspector: Li Haihong, Xiang Chunfang
Aihua

Verifier: Xie