

HUIAN WANJIA MEDICAL DEVICE CO.,LTD.

TEL:0086-517-84519182

FAX:0086-517-83765785

ANALYSIS REPORTS OF STERILE BLOOD LANCET

Merchandise name	Sterile BLOOD LANCET 30G (200PCS/BOX,100PCS/BOX)	Standard
PO NO.:E-19225	LOT:20190706	EN ISO10993;Sampling method:GB/T2828
Item of specification	Standard require	Conclusion of inspection
Hemolysis	Hemolysis rate should be less than 5%	Qualified
Acute Systemic Toxicity	Should have no acute systemic toxicity	Qualified
Cytotoxicity	Cell toxicity level should not greater than 1	Qualified
Sensitization	Skin reactions should not great than 1	Qualified
Intradermal stimulation	Intradermal reaction test should comply with the requirements	Qualified
Sterility	should be sterile	Qualified
Conclusion: This Lot of Products is qualified		
QUALITY DIRECTOR: Yongchao Cheng		DATE:NOV 20,2019

