

## Certificate of Analysis

Product Name	Stainless steel Stethoscope	Model	Child
QTY	500 pcs	Lot Number	20240715
MFG.Date	20240715	Test Date	2024.09.15
No.	Test Item & Requirement	Result	Single Verdict
1	The sound sending of stethoscope shall clear.	Conformity	Qualified
2	The earring's elasticity shall feasible, when 2 earplugs have been undrawn to 300mm distance, after comeback, the distortion distance shall less than 10mm .	Conformity	Qualified
3	The earring bounce shall moderate.	Conformity	Qualified
4	Three-way catheter's surface shall slick and clean, without air bubble and distortion disfigurements etc.	Conformity	Qualified
5	Diaphragm shall clean, without nick, pock and besmirch disfigurements etc. Diaphragm shall not become flexible.	Conformity	Conformity
6	All parts of stethoscope shall fixed, not become flexible.	Conformity	Conformity
7	Each parts' appearance shall symmetry, without crack, sunken and plating break off etc. Plating parts' appearance shall brightness, surface roughness degree Ra0.8, without yellow uncovered, pocked, break off and nicks disfigurements etc.	Conformity	Conformity

Inspected by / Date: Li Kejie / 2024.09.15

Approved by / Date: He Ming / 2024.09.15

