



压缩雾化器成品检验报告 (美国)

Manufacturing Date: 2025/12/18
Expiration date: N/A

Inspection report on finished compressor nebulizer

ALESSA

检验依据 Inspection basis: NTLD-OQC-SIP014 压缩雾化器成品检验工艺 (美)									
产品型号 Product model : NB-2PC		报检数量 Quantity declared : 800 PCS			产品批号 Batch number.: 25YF1B0781				
温度 Temperature: 16℃		湿度 Humidity: 66%			检验人 Inspector: [Signature]				
产品编号 Product number.: 252P00001-252P00800		检验数量 Amount of inspection: 20/32 PCS			检验日期 Date of inspection: 2026-1-30				
序号 NO	检验项目 Inspection	检测设备 Testing device	设备编号 Equipment number	检验水平 Inspection level	AQL 值	抽样数 Sampling number	不良数 Number of defectives	判定 Judgement	备注 Remarks
1	电压 120V 的压力 ≥ 30 psi (极限压力大于 40psi 时, 允许卡死 When the voltage is 120V, the pressure is ≥ 30 psi (extreme pressure is >40 psi), it allows to be jammed.	压力流量计 Pressure flowmeter	FF01-25	S-4	C=0	20	0	合格	
2	电压 120V 时的电流在 < 1600 mA When the voltage is 120V, the current is < 1600 mA	交流电频稳压电源 Ac frequency regulator	E601-02	S-4	C=0	20	0	合格	
3	压力在 12 psi 时流量 ≥ 4.0 L/min ; When the pressure is 12 psi, the flow rate is ≥ 4.0 L/min	压力流量计 Pressure flowmeter	FF01-25	S-4	C=0	20	0	合格	
4	测试电压为 108 V 和 132 V 时, 产品应能正常工作, 出雾正常, 无异响; When the test voltage is 108 V and 132 V, the device should be able to work properly with normal foggy and no abnormal noise.	交流电频稳压电源 Ac frequency regulator	E601-02	S-4	C=0	20	0	合格	
5	雾化器整机最大工作噪音 ≤ 65 dB ; Maximum operating noise of nebulizer ≤ 65 dB	分贝仪 Decibelmeter	SN01-03	S-1	2.5	5	0	合格	
6	雾化率 $A \geq$ ml/min ; Atomization rate $A \geq$ ml/min	天平/秒表 balance/stopwatch	FW01-42 WT01-04	S-4	C=0	20	0	合格	

7	配件装配检查: 组装正确 Assembly inspection of parts: correct assembly	目测 Visual inspection	/	S-1	25	5	0	合格
8	杂色点符合要求、划痕符合要求、组装段差、间隙符合要求 Variegated spots meet the requirements, scratches meet the requirements, poor assembly section, gap meet the requirements	目测 Visual inspection	/	I	600	32	0	合格
9	包装检测 Package inspection	目测 Visual inspection	/	I	600	32	0	合格

综合判定synthetical judgement:

合格

核准 Approval

阴美玲

日期Date:

2026.1.30

记录编号.:NTLD-QS-209-F40

d版

2026013001

实施日期: 2022.02.26

随附检测记录 1 份 Attached two inspection records

2026013002

上海煜丰国际贸易有限公司
SHANGHAI YUFEI INTERNATIONAL TRADE CO., LTD

2026013003

朱文军

检测记录
Inspection record

NO: 2026013001

检测序号 Test NO. Serial number 产品编号	1	2	3	4		5	6	7	8	9	判定 judgement	备注
	$\geq 30 \text{ psi}$	$< 1600 \text{ m/s}$	$> 0.01/\text{min.}$	108V	132V	$\leq 65 \text{ dB}$	附件	配件	包装	外观		
2521900036	69.8	1032	6.0	✓	✓	50.3	/	✓	✓	✓	✓	
00074	50.5	1027	5.5	✓	✓	50.1		✓	✓	✓	✓	
00120	52.1	1031	5.5	✓	✓	69.8		✓	✓	✓	✓	
00159	50.3	1032	5.5	✓	✓	50.1		✓	✓	✓	✓	
00197	50.2	1026	5.5	✓	✓	50.1		✓	✓	✓	✓	
00235	50.6	1016	5.5	✓	✓			✓	✓	✓	✓	
00281	50.5	1018	5.5	✓	✓			✓	✓	✓	✓	
00323	52.1	1069	5.5	✓	✓			✓	✓	✓	✓	
00360	50.9	1051	5.5	✓	✓			✓	✓	✓	✓	
00600	51.0	1068	5.5	✓	✓			✓	✓	✓	✓	
00641	53.4	1066	5.5	✓	✓			✓	✓	✓	✓	
00682	51.7	1065	6.0	✓	✓			✓	✓	✓	✓	
00523	50.6	1066	5.5	✓	✓			✓	✓	✓	✓	
00558	50.2	1053	5.5	✓	✓			✓	✓	✓	✓	
00596	50.4	1069	5.5	✓	✓			✓	✓	✓	✓	
00639	69.3	1035	5.5	✓	✓			✓	✓	✓	✓	
00680	51.8	1052	5.5	✓	✓			✓	✓	✓	✓	
00720	69.3	1037	5.5	✓	✓			✓	✓	✓	✓	
00760	68.2	1038	5.5	✓	✓			✓	✓	✓	✓	
00797	69.7	1066	5.5	✓	✓			✓	✓	✓	✓	
00021							✓	✓	✓	✓		
0053							✓	✓	✓	✓		

2026.1.30

