



北京市津威康达医疗器械有限公司
Beijing Jinwei Kangda Medical Instrument Ltd.

Certificate of Analysis

Name	Knitted Knee brace II		Specification	L
The quantity of sample	4 Pieces		Lot No.	24051001
Model	KDHX-13		Production Date	20240510
Inspection Reference	JINGFENGXIEBEI20150036		Test item	Product registration standard
Test results	Item	Standard	Actual results	Judgment for each
	Appearance	Should be no dirty ,no thread splitting, and no damaged , the both size of support bar and enhance belt are protected well.	Meet the standard	Qualified
	Lateral elasticity test	1.2-1.7 times		Qualified
	Metal residue detection	Should pass the testing of table needle detection machine		Qualified
	Test condition	In the condition of 20-30 ℃ temperature and 50-70% humidity, test all the basic parameters, take the avarage value of three times as test result.		
Inspection Conclusion	According to the test above all the test items are qualified.			

Checked by: Donglin Ge

Checked date: 2024/06/20

Company: Beijing Jinwei Kangda Medical Instrument Ltd.

