

FOSHAN FEIYANG MEDICAL EQUIPMENT CO., LTD.
CERTIFICATE OF ANALYSIS

NO: FSFY-COA20250823005D

Product Name	SHOWER CHAIR / SILLA DE DUCHA	Lot.No.	FY20250823	QTY	300PCS	Test QTY	300PCS	Standard	GB/T 24434 - 2009 (ISO13485: 2016)
Model No.	FY793L	Inspector	Liu zhenyan	MFG Date	Aug.23,2025	checker	PENG LI	DATE	Aug.23,2025
NO.	Name of item	Technology specification						Test Result	Conclusion
1	Appearance structure	Appearance should be smooth, without irregularity , scars, distortion or deflection .						Compliant	Pass
2	Dimension	Base size of shower chair should match the requirement of Form 1, limit deflection is no more than ±5cm.						Compliant	Pass
3	Stability	Four legs should be stable on the floor without swing.						Compliant	Pass
4	Colour transfer	Plastic parts should have no colour transfer.						Compliant	Pass
Test result	According to the test, all qualification reach the manufacture's standard request GB/T 24434 - 2009(ISO13485: 2016) . Quality managing department: PAN QIU WEI Date: Aug.23,2025								

