

**FOSHAN FEIYANG MEDICAL EQUIPMENT CO., LTD.**  
**CERTIFICATE OF ANALYSIS**

NO: FSYF-COA20250825002A

Product Name	STICK / BASTON	Lot.No.	FY20250825	QTY	1200PCS	Test QTY	1200PCS	Standard	GB/T 19545.1-2012 (ISO13485: 2016)
Model No.	FY921	Inspector	PENG HONGLIN	MFG Date	Aug.25,2025	checker	PENG LI	DATE	Aug.25,2025
<b>NO.</b>	<b>Name of item</b>	<b>Technology specification</b>						<b>Result</b>	<b>Single conclusion</b>
1	Appearance structure requirement	Appearance should be smooth, without irregularity , scars, distortion or deflection .						Compliant	Pass
2	Dimension	Base size of stick should match the requirement of Form 1, limit deflection is no more than $\pm 5$ cm.						Compliant	Pass
3	Stability	The leg should be stable on the floor without swing.						Compliant	Pass
4	Colour transfer	Plastic parts should have no colour transfer.						Compliant	Pass
Test result	According to the test, all qualification reach the manufacture's standard request GB/T 19545.1-2012 (ISO13485: 2016)								
Quality managing department: PAN QIU WEI								Date:Aug.25,2025	

